

EC-5385-CS5 EN-1948-4 Marker PCB Calibration Series CS5 (unlabeled/¹³C₁₂, 99%) 0.2 mL
in Isooctane:Nonane

Native Compounds	IUPAC#	Concentration (ng/mL)
2,4,4'-TriCB	28	500
2,2',5,5'-TetraCB	52	500
2,2',4,5,5'-PentaCB	101	500
2,2',4,4',5,5'-HexaCB	153	500
2,2',3,4,4',5'-HexaCB	138	500
2,2',3,4,4',5,5'-HeptaCB	180	500
¹³C-Labeled Sampling Standards	IUPAC#	Concentration (ng/mL)
2,3,4,4'-TetraCB (¹³ C ₁₂ , 99%)	60	10
3,3',4,5,5'-PentaCB (¹³ C ₁₂ , 99%)	127	10
2,3,3',4,5,5'-HexaCB (¹³ C ₁₂ , 99%)	159	10
¹³C-Labeled Extraction Standards	IUPAC#	Concentration (ng/mL)
2,4,4'-TriCB (¹³ C ₁₂ , 99%)	28	100
2,2',5,5'-TetraCB (¹³ C ₁₂ , 99%)	52	100
2,2',4,5,5'-PentaCB (¹³ C ₁₂ , 99%)	101	100
2,2',4,4',5,5'-HexaCB (¹³ C ₁₂ , 99%)	153	100
2,2',3,4,4',5'-HexaCB (¹³ C ₁₂ , 99%)	138	100
2,2',3,4,4',5,5'-HeptaCB (¹³ C ₁₂ , 99%)	180	100
¹³C-Labeled Recovery Standards	IUPAC#	Concentration (ng/mL)
2,3',4',5-TetraCB (¹³ C ₁₂ , 99%)	70	10
2,3,3',5,5'-PentaCB (¹³ C ₁₂ , 99%)	111	10
2,2',3,3',4,4',5-HeptaCB (¹³ C ₁₂ , 99%)	170	10